

CFM Olympic Brainz Monitor

complements neonatal bedside brain monitoring with automatic marking of possible seizure activity.

RecogniZe the problem

FACT: Neonatal seizures are usually clinically subtle, inconspicuous and difficult to recognize from the normal behaviors of the inter-ictal periods or physiological phenomena¹, making accurate recognition and treatment challenging for clinicians in the busy NICU environment.

FACT: Seizures are more common in the neonatal period than in any other time in life – as high as 57.5/1000 in < 1500g and 2.8/1000 in 2500-3999g.²

FACT: With moderate-to-severe hypoxic-ischemic encephalopathy (HIE), the incidence of seizures is > 50%.³


FACT: Term infants with seizures have very poor outcomes, with a 20% fatality rate in the neonatal period. Survivors have a 28% - 35% risk for severe neurodevelopmental disability and a 20% - 50% risk for epilepsy.⁴

RecogniZe the hospital challenge

FACT: In many institutions there is a lack of round-the-clock neurological resources to identify at-risk infants, increasing the potential for poor outcomes.

FACT: Conventional EEG (cEEG) using greater than 10 electrodes and concurrent video recording is the gold standard for bedside monitoring. However, cEEG is labor-intensive to both set up and interpret, and interpretation is rarely available in real time.⁴

FACT: Unidentified and untreated neonatal seizures can lead to long-term impairment or death.¹



RecogniZe the solution

The CFM Olympic Brainz Monitor's RecogniZe option assists clinicians in identifying possible seizure activity:

- Developed for the NICU for ease of use by NICU staff – reducing the need to bring in neurology staff and equipment
- Helps ease identification and validation of seizure activity in the raw EEG signal

Monitoring for seizures with limited-channel amplitude-integrated EEG (aEEG) can be accurately interpreted. It also compares favorably with cEEG, and aligns with a general trend toward reduced seizure burden.⁴

It has been shown that the anatomical regions conventionally used for aEEG electrode placement yield a high sensitivity (78%) for electrical seizure detection.⁵

Ordering Information

Item	Part number
OBM RecogniZe License Kit	OBM00092

References:

- 1 Panayiotopoulos CP. The Epilepsies: Seizures, Syndromes and Management. Oxfordshire (UK): Bladon Medical Publishing; 2005.
- 2 Volpe JJ. Neonatal Seizures, Chapter 5. Neurology of the Newborn, 4th edition. W. B. Saunders Company, 2001.
- 3 Gluckman PD, Wyatt JS, Azzopardi D, Ballard R, Adwards AD, Ferriero DM, Polin RA, Robertson CM, Thoresen M, Whitelaw A, Gunn AJ. Selective head cooling with mild systemic hypothermia after neonatal encephalopathy: multicentre randomized trial. The Lancet. 2005;365:663-70.
- 4 Lawrence R, Mathur A, Nguyen The Tich S, Zempel J, Inder T. The Journal of Pediatrics. 2009;154(6):835-841.e1.
- 5 Shellhaas RA, Soaita AI, Clancy RR. Sensitivity of Amplitude-Integrated Electroencephalography for Neonatal Seizure Detection. Pediatrics. 2007;120:770-777.

natus
Natus Medical Incorporated
1501 Industrial Road
San Carlos, CA 94070 USA
1-800-303-0306
+1-650-802-0400
www.natus.com

© 2013 Natus Medical Incorporated. All Rights Reserved. All product names appearing on this document are trademarks or registered trademarks owned, licensed to, promoted or distributed by Natus Medical Incorporated, its subsidiaries or affiliates.



A trend is growing in many health care facilities to adopt the CFM Olympic Brainz Monitor with the sole intention of improving their NICU staff's ability to identify sub clinical seizures.

