

**Type identification**

MADSEN Zodiac Quick Check is type 1096 from Natus Medical Denmark ApS

**Feature table**

	Quick Check	Diagnostic	Clinical
Tympanometry, auto	X	X	X
Tympanometry, manual		X	X
Tymp History	X	X	X
Probe tone, 226 Hz	X	X	X
Probe tone, 1000 Hz		X	X
Probe tone, 678 and 800 Hz			X
Reflex Screening	X	X	X
Reflex Threshold (Ipsi/Contra)		X	X
Reflex Decay		X	X
ETF-I (Intact)		X	X
ETF-P (Perforated)			X
B & G tympanograms			X
Admittance Recorder			X

**Compliance measuring system**

Probe tone:	226 Hz at 85 dB SPL $\pm$ 3 dB
Dynamic probe tone level:	The probe tone level will be compensated to accommodate varying ear canal volumes. The output level will be decreased in volumes < 1.7 ml. The output level will be increased in volumes > 2.3 ml
THD:	< 1% in 2 cc
Frequency accuracy:	$\pm$ 0.5%
Range:	0.2 ml to 5.0 ml $\pm$ 5% or 0.05 ml whichever is greater * 5.0 ml to 8.0 ml $\pm$ 15% * * The accuracy stated requires that calibration has been performed at the altitude where the device is to be used.

**Acoustic reflex****Sensitivity**

Reflex Screening:	0.04 mmho
Step size dB:	Quick Check: 10 dB

**Ipsilateral Stimulation**

Tone:	500 Hz, 1000 Hz, 2000 Hz, 4000 Hz
Frequency accuracy:	$\pm$ 0.5%
Range:	BBN, LPN, HPN at 50 to 110 dB SPL * $\pm$ 3 dB (* measured in calibration coupler)
Screening range:	BBN at 50 to 90 dB SPL * $\pm$ 3 dB (* measured in calibration coupler)

**Air pressure system**

Range:	Normal +200 to -400 daPa/s
Pressure sweep rate:	200, 400, 600 daPa/s $\pm$ 20% in 20% to 80% of the total pressure range
Pressure accuracy:	$\pm$ 10% or $\pm$ 10 daPa, whichever is greatest
Pump measure direction:	Positive to negative or negative to positive
Safety:	Separate safety +530 daPa and -730 daPa $\pm$ 70 daPa

**Graph units**

Unit of admittance graph Y-axis:	ml, cc, mmho, $\mu$ l
Unit of graph X-axis:	daPa, sec

**Device display**

Display:	7 inch, 15:9 WVGA
Resolution:	800 x 480 pixel

**USB port connector**

Type:	USB device port
-------	-----------------

**Power supply**

External power supply	XP Power, type AFM60US24
Output:	24 V, 2.5 A
Input:	100-240 V AC, 50-60 Hz, 1.5 A

**Power consumption**

Power consumption:	< 70 VA
--------------------	---------

**Operating environment**

Temperature:	+15°C to +35°C (59°F to +95°F)
Air humidity:	10 to 90%, non-condensing
Air pressure:	60 kPa to 106 kPa
Warm-up time:	< 10 min. If stored in conditions not within specified operating environment conditions, the device must warm up for 24 hour before being used.

**Storage and handling**

Temperature:	-20°C to +60°C (-4°F to +140°F)
Relative humidity:	< 90 %, non-condensing
Air pressure:	50 kPa to 106 kPa

**Dimensions (HxWxD)**

Stand-alone version:	190 mm x 248 mm x 261 mm (7.5" x 9.8" x 10.3")
PC-based version:	100 mm x 240 mm x 240 mm (3.9" x 9.4" x 9.4")

**Probe dimensions (HxWxD)**

Quick Check probe:	28 mm x 22 mm x 125 mm (1.1" x 0.9" x 4.9"), Cable length 2800 mm (110")
--------------------	--

**Weight**

Stand-alone version:	2.65 kg/5.85 lb
PC-based version:	1.65 kg/3.64 lb

**Optional features (Stand-alone)**

Printer:	Built-in printer. Prints 832 dot line/s on 112 mm paper width
2 cc coupler	

**Calibration**

Equipment should be calibrated regularly according to EN 60645-5 and ANSI S3.39

**Standards**

Safety:	IEC 60601-1:2005+AMD1:2012 EN 60601-1:2006+A1:2013 ANSI/AAMI ES60601-1:2005 + A1:2012 CAN/CSA-C22.2 NO. 60601-1:14 Class II, externally powered, Type BF, IPX0
EMC:	IEC 60601-1-2:2007 and EN 60601-1-2:2007 IEC 60601-1-2:2014 and EN 60601-1-2:2015
Impedance/Admittance:	Quick Check: EN 60645-5:2005 Type 2, ANSI S3.39 1987 (R2012) Type 2
Power supply:	Class I, externally powered supply

**System requirements**

For system requirements, please refer to the Otosuite data sheet.

*Specifications are subject to change without notice. Copyright © Natus Medical Denmark ApS. Doc. No. 7-26-4704-EN/05 - Part No. 7-26-47004-EN (214038)*